



MISSOURI DEPARTMENT OF HEALTH
STATE PUBLIC HEALTH LABORATORY
DATAMASTER MAINTENANCE REPORT

RECEIVED On-Site Rapid Response
By Carol Day at 12:29 pm, Mar 09, 2010

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send copy to Department of Health; retain original in department file.

DATAMASTER SN 940374	DATE OF INSPECTION 03/02/2010
LOCATION OF INSTRUMENT (STREET AND CITY) 201 North Street, Waynesville, MO 65583	TIME OF INSPECTION 09:37

CHECKLIST: Place a check (✓) to the left of each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unchecked items must be corrected before using instrument.

☒ DIAGNOSTIC CHECK (PRINTOUT ATTACHED)

☒ COMPUTER

☒ DETECTOR

☒ PROGRAM

☒ FILTERS

☒ HEATERS SAMPLE CHAMBER **49** °C

☒ QUARTZ STANDARD

☒ FLOW DETECTOR

☒ CALIBRATION

☒ PUMP HIGH SPEED

☒ PRINTER

☒ INDICATOR LIGHTS

☒ TIME AND DATE

☒ SIMULATOR TEMPERATURE (34°C ± 0.2°C) **34** °C

☒ CALIBRATION CHECK -
Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
(USE) RECIRCULATION PUMP

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

TEST 1 ◀ **.101**

TEST 2 ◀ **.101**

TEST 3 ◀ **.101**

☒ PERFORM RFI TEST (PRINTOUT ATTACHED)

☒ NUMBER OF REFUSALS, SINCE LAST MAINTENANCE REPORT, AND NUMBER OF BREATH TESTS IN EACH RANGE AS FOLLOWS: (DO NOT INCLUDE SIMULATOR TESTS)

REFUSALS	0	(0 -.04)	0	(.05 -.09)	0	(.10 -.14)	0	(.15 -.19)	0	(Over .19)	1
----------	----------	----------	----------	------------	----------	------------	----------	------------	----------	------------	----------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary)

THIS INSTRUMENT OPERATES WITHIN DEPARTMENT OF HEALTH STANDARDS.

SUPPLIER: REPCO MARKETING INC.

CONCENTRATION: .10

EXPIRATION: 04/22/2011

LOT NUMBER: 09001

INSPECTING OFFICER

SIGNATURE

PRINT NAME

Assistant Chief of Police Clarence Liberty

TYPE II PERMIT NUMBER / EXPIRATION DATE

TELEPHONE NUMBER

820337/12-22-2010

573-774-2414/573-774-2198

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 09001
EXPIRATION DATE: April 22, 2011 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 09001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1214 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is April 23, 2009.
The expiration date for this lot number is April 22, 2011
at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



CLARENCE H LIBERTY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 12/22/2008

Number 820337

Expires 12/22/2010

MO 580-0771 (7-88)

John J. Mathewson

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
WAYNESVILLE POLICE DEPARTMENT

SAC DATAMASTER SERIAL NUMBER: 949374
03-02-10

TESTING OFFICER:
LIBERTY/CLARENCE
OFFICER I.D.# 701
PERMIT NUMBER: 020337
EXPIRATION DATE: 12-22-10
MISCELLANEOUS DATA:
CALIBRATION CHECK

- SUPERVISOR NOTE -

BLANK TEST	.000	09:46
INTERNAL STANDARD	VERIFIED	09:47
EXTERNAL STANDARD	.101	09:47
BLANK TEST	.000	09:48
EXTERNAL STANDARD	.101	09:48
BLANK TEST	.000	09:49
EXTERNAL STANDARD	.101	09:49
BLANK TEST	.000	09:50

N = 3
SIN. = .1
AVG. = .101

Operator Signature

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
WAYNESVILLE POLICE DEPARTMENT

SAC DATAMASTER SERIAL NUMBER: 949374
03-02-10
09:37

- DIAGNOSTIC CHECK -

COMPUTER:	OKAY
PROGRAM (04-07-2000):	OKAY
WEIGHTS	
SAMPLE CHAMBER:	45L
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ;
@ A B C D E F G H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ! " # \$ % ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ;

Operator Signature

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
WAYNEVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940374
03/02/10


ARREST TIME: 09:30
SUBJECT NAME:
TIME/BLANK/TEST
DOB: 09/00/70 SEX: M
STATE/D.L.: MO-000000
ARRESTING OFFICER:
LIBERTY/CLARENCE
OFFICER I.D.: 701
TESTING OFFICER:
LIBERTY/CLARENCE
OFFICER I.D.: 701
PERMIT NUMBER: 000007
EXPIRATION DATE: 12/22/18
MISCELLANEOUS DATA:
TIMED BLANK BREATH TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	09:57
INTERNAL STANDARD	VERIFIED	09:50
SUBJECT SAMPLE	.000	09:50
BLANK TEST	.000	10:00

15 seconds

Operator Signature



Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
WAYNEVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940374
03/02/10

ARREST TIME: 09:30
SUBJECT NAME:
RFI/TEST
DOB: 09/00/70 SEX: M
STATE/D.L.: MO-000000
ARRESTING OFFICER:
LIBERTY/CLARENCE
OFFICER I.D.: 701
TESTING OFFICER:
LIBERTY/CLARENCE
OFFICER I.D.: 701
PERMIT NUMBER: 000007
EXPIRATION DATE: 12/22/18
MISCELLANEOUS DATA:
AUDIO INTERFERENCE TEST

--- BREATH ANALYSIS ---

AUDIO INTERFERENCE

Operator Signature

